KWAZULU-NATAL PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

ANNEXURE A: SPECIFICATION FORM

NAME OF PROCURING FACILITY	VRYHEID HOSPITAL PHARMACY	
ITEM DESCRIPTION	ELECTRONIC VACCINE STORAGE REFRIGERATOR WITH ICE PACK FREEZER	
ITEM PURPOSE	1. VACCINE STORAGE 2.	
ITEM DETAILED SPECIFICATION (INCLUDE SIZE, COLOUR, MATERIAL, ETC.)		COMPLIES (YES/NO)
1.	International Cold Chain Compliant – WHO APPROVED	
2.	SAHPRA APPROVED - WITH 6 SHELVES	
3.	DUAL COOLING SYSTEM (SEPARATE ON/OFF SWITCH)	
4.	GROSS REFRIDGERATOR VOLUME = 288L, FREEZER = 48L	
QUALITY STANDARD	WHO, SAHPRA	
UNIT OF MEASURE OR PACKAGING I.E. (UNIT/BOX/ROLL/PACK/BAIL ETC)	2	
SAMPLE REQUIRED (YES/NO) IF YES WHEN AND HOW?	NO	
ADDENDUM TO SPECIFICATION ATTACHED (YES OR NO)	Yes	

Note: MUST HAVE !-

1. SOLID DOOR FINISH
2. TEMPERATURE (FRIDGE) + 2°C to + 8°C at 35°C AMBIENT
3. ACCURATE TEMPERATURE CONTROL, INTERIOR LIGHT WITH STAINLESS STEEL INNER LINER.
4. DOOR LOCKING SYSTEM
5. AUTOMATIC DEFROST
6. HOLD OVER TIME DURING POWER CUT 7 48 HOURS
7. DIMENSIONS (mm) H = 1830mm X W = 820mm X D = 710 mm