Specification:

1.1 The ISO 9000 or ISO 13485 certificate of manufacturer shall be attached

1.3 A signed letter from the registered manufacturer (on the Manufacturer letterhead) supporting and underwriting the vendor as the only accredited and certificated vendor in South Africa for safe distribution, service and maintenance

1.4 Minimum of 2-year warranty should be included in the basic price. Attach manufacturer proof in writing.

1.5 The colour barouche must be attached

Technical Specification:

2.1 A durable medical utility carts which can be used for patient files is required.

2.2 The unit to have an ABS work-top with raised edges to prevent items from falling

2.3 The unit to be constructed from the

2.4 Lightweight aluminium frame or equivalent corrosion-resistant material

2.5 The trolley to be equipped with Integrated ergonomic push handles

2.6 The trolley to have Minimum three lockable drawers comprising: 2 × shallow drawers (approximately 75 mm/3") and 1 × medium drawer (approximately 150 mm/6")

2.7 Central locking system with key lock securing all drawers simultaneously

2.9 The overall dimensions of Approximately 760 mm (W) × 490 mm (D) × 1030 mm (H) (+- 10mm)

