KWAZULU-NATAL PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

ANNEXURE A: SPECIFICATION FORM

NAME OF PROCURING FACILITY	Mosons, & Matuddle Scottsville
ITEM DESCRIPTION	Thermal blanket
ITEM PURPOSE	Use during resuscitation of hypothermia cases

ITEM DETAILED SPECIFICATION (INCLUDE SIZE, COLOUR, MATERIAL, ETC.)		COMPLIES (YES/NO)
1.	Aluminium foil laminated or plastic sheeting	
2.	Silver 2.1m X 1.3m	
3.	25 packs in the box	
4.		
5.		
QUALITY STANDARD		
UNIT OF MEASURE OR PACKAGING I.E. (UNIT/BOX/ROLL/PACK/BAIL ETC)	BOX	
SAMPLE REQUIRED (YES/NO) IF YES WHEN AND HOW?	NO	
ADDENDUM TO SPECIFICATION ATTACHED (YES OR NO)	NO	

Note:

1.
2.
3.

SPECIFICATION APPROVED BY

Name of End-user (in full)	~~Diane Carson~~ S.S Nxumalo	Name of SCM Rep (in full)	S. Khamula.
Designation / Rank (in full)	Operational Manager	Designation/ Rank (in full)	SCC.
Signature		Signature	
Date	03.08.26	Date	03.08.26

Bidder Initial here: _____

ANNEXURE A: SPECIFICATION FORM

NAME OF PROCURING FACILITY	**Masons Clinic**	
ITEM DESCRIPTION	Casualty stitch pack	
ITEM PURPOSE	1.	
ITEM DETAILED SPECIFICATION (INCLUDE SIZE, COLOUR, MATERIAL, ETC.)		**COMPLIES (YES/NO)**
1.	2. 16 piece including tray	
3.	Surgical grade Stainless Steel (304→420)	
4.	To withstand repeated autoclave Sterilisation	
5.		
6.		
QUALITY STANDARD	ISO 13485	
UNIT OF MEASURE OR PACKAGING i.E. (UNIT/BOX/ROLL/PACK/BAIL ETC)	**Pack**	
SAMPLE REQUIRED (YES/NO) IF YES WHEN AND HOW?	NO	
ADDENDUM TO SPECIFICATION ATTACHED (YES OR NO)	**Yes**	

Note: To contain the following

1. Needle holder
2. Scalpel handle #3 and #4
   Scissors - Straight / operating
   Forceps - Mosquito - Straight + Curved
   Dressing
   Tissue
   Lotion bowl
   gallipot
   Stainless steel instrument tray

SPECIFICATION APPROVED BY

Name of End-user (in full)	Diane Carson	Name of SCM Rep (in full)	S. Knapyle.
Designation / Rank (in full)	Operational Manager	Designation/ Rank (in full)	29.07.26 SCC.
Signature		Signature	
Date	01/06/2026	Date	29.07.26

Bidder Initial here: _____

KWAZULU-NATAL PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

ANNEXURE A: SPECIFICATION FORM

NAME OF PROCURING FACILITY	Masons Clinic	
ITEM DESCRIPTION	Aseptor self-sealing pouches	
ITEM PURPOSE	1. For Sterilising light weight instruments	
ITEM DETAILED SPECIFICATION (INCLUDE SIZE, COLOUR, MATERIAL, ETC.)		COMPLIES (YES/NO)
1.	200 / Box	
2.	Medical-grade paper back / transparent front.	
3.	90mm X 260mm "Size"	
4.		
5.		
QUALITY STANDARD	ISO 13485	
UNIT OF MEASURE OR PACKAGING I.E. (UNIT/BOX/ROLL/PACK/BAIL ETC)	Box of 200	
SAMPLE REQUIRED (YES/NO) IF YES WHEN AND HOW?	No ✓	
ADDENDUM TO SPECIFICATION ATTACHED (YES OR NO)	NO	

Note:

1.



SPECIFICATION APPROVED BY

Name of End-user (in full)	Diane Carson	Name of SCM Rep (in full)	S. Ramlio.
Designation / Rank (in full)	Operational Manager	Designation/ Rank (in full)	SCC.
Signature		Signature	
Date	01/06/2026	Date	03.06.26

Bidder Initial here: _____