593 Potgieter Rd, Glen Austin Midrand
Johannesburg
Contact: 083 641 3776
Email: mphatsoejohn@gmail.com

Date: 21 May 2026

REQUEST FOR A QUOTATION

Description	Unit of Measure	Quantity
Citadel Bed plus Bariatric 350kg **Bed Citadel plus Bariatric 350kg** Bed, bariatric, ICU, height of the mattress support shall be adjustable in the range 400 – 780 mm (measured from the floor to the top of the mattress support), with tolerance of +40 mm, backrest angle: 65o (+5o), knee-break angle: 35o (+5o), trendelenburg angle: 15o (+3o), reverse trendelenburg angle: 15o (+3o), maximum patient weight shall not be less than 350kg, overall dimensions: length: 1 980 mm (+40mm), width: 900 mm (+20mm), thickness: 150 mm (+5mm), electrically operated variable height adjustment, backrest, knee-break, reverse trendelenburg and trendelenburg, safe working load shall not be less than 450kg, must have a battery back-up in case of power failure, to be fitted with castors with a central castor locking system, must be fitted with Oxygen cylinder holder	**EACH**	**1**

Regards

John Mphatsoe
Director
MIG